





UK NEQAS H&I Annual Participant's Meeting 2024-25







Prifysgol Felindre Velindre University



Meet The Team!

Director: Deborah Pritchard Manager: Amy De'Ath **Deputy Manager**: Melanie Bartley Healthcare Scientist Practitioner: Geraint Clarke **QA Technical Officer**: Jack Jefferies MLA: Sue Davies



UKNEQASHandl@Wales.NHS.UK



UK NEQAS for H&I Steering Committee 2024-25

Helena Lee (Chair) Arthi Anand Katy Derbyshire Sylvia McConnell Katherine Mounsey Anthony Calvert Sunil Daga (Clinical Representative) Elizabeth Wroe (BSHI Representative to UK NQAAP)

> Rhys Goodhead (Expert Advisor Scheme 5B) Barbara McNamara (Expert Advisor Scheme 5B) Tim Clench (Expert Advisor Scheme 5B)

UK NEQAS for H&I: An Overview

>320 participants



>50 countries



UK NEQAS for H&I is 50!

1975	 National Tissue Typing and Reference Laboratory (now OTDT) in Bristol initiated a quality control scheme for Exercises included technical comparisons e.g. comparing batches of complement and the sensitivity of difference 	HLA typing (1A) and crossmatching (2A) rent techniques
1988	Introduction of HLA antibody specification (Scheme 3)	BI: TRANSPLAT EXTINCE Arrow A Analogy
1989	First Annual Participant Meeting Idea for a UK professional society for tissue typers: BSHI formed	The Quality Assessment Every in The Second State
1990	• HLA disease association scheme – B27	The report of last years for the second seco
1992	• HLA genotyping scheme introduced: Class II only (Class I not introduced until 1999)	b) Sufficient and
1994	First international participants joined	contrast enti-ref.at class mon. Lines 4 propriated summary, in britch, such presents and the second state of the second state
1998	Founding member of the EFI EPT Committee	on the typing results to privat for early in,

l	JK	NEQAS for H&I is 50!	
	2000	• UK NEQAS for H&I relocated to the Welsh Blood Service • New HFE (5) and antibody detection (6) schemes introduced • First educational schemes for HLA typing	
	2008	• Drug hypersensitivities introduced for Abacavir (7)	50 years of UK NEGAS
	2011	• 2 nd Field HLA Genotyping (4A2) developed • HLA disease association (8) and KIR genotyping (9) introduced	
	2015	ISO 17043 accreditation HPA genotyping pilot (10)	
	2016	• HPA Antibody pilot (11)	
	2018	• 3 rd Field HLA Genotyping	
5	2025	• HNA genotyping (12) and HNA antibody detection (13)	

De'Ath A, Rees MT, Pritchard D. The history and evolution of HLA typing external proficiency testing schemes in UK NEQAS for H&I. Front Genet. 2023 Sep 18;14:1272618. doi: 10.3389/fgene.2023.1272618. PMID: 37790700; PMCID: PMC10544324.

(



Things To Note...



Presentation Focus... Performance, key trends, discussion points and 2025 changes



Further Details...

The presentation will be available to view on our website.



Lab Locations... Generally: 1-100 = UK & Ireland. 101+ = Rest of the world



Scheme Assessments



- Most Schemes assessed on a consensus basis using a 75% consensus level i.e. 75% of reports must agree on a result for it to be assessed.
- Reference typing results are used for typing/disease schemes if consensus not reached plus educational schemes where required:
- e.g. Scheme 8: HLA Genotyping for Coeliac and Other HLA Associated Diseases
- Equivocal result only accepted for Scheme 2B.
- All Not Tested (NT) results excluded from assessment.
- Labs that fail to return results or do not a provide valid reason for NT are assessed as unacceptable.

Unsatisfactory Performance (UP)

• Each scheme has minimum annual performance criteria:

- ► HLA Typing schemes 90%
- ► Crossmatching 85%
- ► Disease Association Schemes 100%
- Antibody Specificity 75%
- Antibody Detection 80%



- Participants that do not meet the minimum criteria are classed as unsatisfactory performers.
- Must complete a root cause analysis and CAPA form.





Changes for 2025-26

Pilot Schemes

Scheme 12 – HNA Genotyping Scheme 13 – HNA Antibody Detection

Sustainability

Paperless: no distribution slips Packaging: reduce plastic



Courier

Reduction in costs Used for domestic and international

Scheme Changes

Scheme 7 – HLA-related Pharmacogenetics Scheme 8 – now only HLA Associated Diseases

Pilot Schemes





Assess participants ability to correctly determine HNA polymorphisms



4 blood samples in two distributions



Pilot Schemes are free of charge and not formally assessed





Assess participants ability to correctly detect the presence and specificity of HNA antibodies



4 serum/plasma samples in two distributions



Pilot Schemes are free of charge and not formally assessed





Cytotoxic Crossmatching



Scheme 2A – Cytotoxic Crossmatch



10 blood samples, 40 serum samples over 5 distributions



Scheme 2A: Performance

All cells with and without DTT	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	71	71	66	63	59	47	47
	(18)	(22)	(16)	(15)	(15)	(10)	(8)
Number with Unsatisfactory Performance (< 85%) (UK&I)	16 (7)	5 (1)	7 (0)	4 (D)	6 <mark>(3)</mark>	2 (2)	5 (0)
% Unsatisfactory Performance	22.5%	7.0%	10.6%	6.3%	10.2%	4.2%	10.8%
(UK&I)	(38.8%)	(4.5%)	(0)	(0)	(20%)	(20%)	(0%)

2024: 5 Unsatisfactory Performers (0 UK & Ireland)



Scheme 2A: Performance by category

	PBL	PBL +DTT	T Cell	T Cell +DTT	B Cell	B Cell +DTT
Crossmatches assessed (n=40) (UK&I)	31 (31)	33 (33)	36 (39)	34 (36)	25 (33)	30 (36)
NT – Assessed samples only	17.1%	14.2%	37.8%	33.5%	56.9% 1	53.6%
% incorrect assignments	4.1%	3.0%	9.3%	7.1%	20.4%	11.3%
False Positive	3.2%	1.5%	5.5%	5.3%	11.2%	5.0%
False Negative	0.9%	1.5%	3.8%	1.8%	9.2%	6.3%

Scheme 2A: Unacceptable Performers 2024



Lab ID	PBL -DTT	T -DTT	B -DTT	PBL + DTT	T + DTT	B + DTT	Lab Identified Error
133		80%					No response
189	81.3%			82.4%			Procedural error
226			84.4%				Interpretation criteria (pos cut off)
232						83.3%	Sample quality (delivery delays)
1412	75%	74.4%					Poor cell prep / reagent issues

Scheme 2A: Discussion



- Not all Scheme 2A results will reach consensus (that's ok!)
- B-cells are difficult (transport, non-specific binding)
- Only partially emulates clinical practice
- 2A is a technical assessment of cytotoxic crossmatching and should not be 'interpreted'
- Lab's need to ensure that all test parameters and acceptance criteria are met prior to reporting NEQAS samples
 - CDC assays are not quantitative so reliant on subjective assessment





Crossmatching by Flow Cytometry

Scheme 2B: Crossmatching by Flow Cytometry



10 blood samples, 40 serum samples over 5 distributions

Q

Scheme 2B: Performance

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UKEI)	83	84	80	80	84	83	79
	(22)	(23)	(21)	(22)	(19)	(21)	(20)
Number with Unsatisfactory Performance (< 85%) (UKBI)	15 (2)	12 (1)	11 (0)	5 (0)	6 (2)	10 (0)	4 (0)
% Unsatisfactory Performance	18.1%	14.2%	13.8%	6.3%	7.1%	12%	5%
(UK&I)	(9.1%)	(4.3%)	(0)	(0)	(10.5%)	(0)	(0)

2024: 4 Unsatisfactory Performers (0 UK & Ireland)



Scheme 2B: Performance by Category

			T Cells			B Cells	
		UK&I	RoW PC	RoW WB	UK&I	RoW PC	RoW WB
	Number of participants	20	28	26	20	28	24
-	Number of XM assessed (>75% consensus)	35/40 (87.5%)	31/40 (77.5%)	36/40 (90%)	36/40 (90%)	36/40 (90%)	36/40 (90%)
	Number of Positive XM	24 (69%)	20 (65%)	22 (61%)	32 (89%)	29 (81%)	28 (78%)
5	Number of Negative XM	11 (31%)	11 (35%)	14 (39%)	4 (11%)	7 (19%)	8 (22%)
	Number of incorrect assignments	20	37	21	18	41	36
	Number of False Pos	10	12	13	6	14	12
	Number of False Neg	10	25	8	12	27	24
	Number of equivocal assignments	0	6	0	1	9	1
	Number of samples NT	25	162	58	56	111	64

UK&I and RoW receive different blood samples



Scheme 2B: Unacceptable Performers 2024

Lab	T Cell	No. of results submitted	B Cell	No. of results submitted	Issue
118	90%	24/40	77%	24/40	Sample quality (delivery delay)
139	87%	26/40	78%	20/40	No response
189	42%	32/40	41%	32/40	Technical issue (cytometer)
191	97%	40/40	83%	40/40	Sensitivity issue (delivery delays/poor cell prep)







HLA Antibody Detection



Scheme 6: HLA Antibody Detection

AND'

Purpose Assess participants ability to determine presence or absence of HLA antibodies

Satisfactory Performance 80% reports agree with consensus in distribution year Consensus At least 75% agreement on presence/absence of HLA antibodies

12 serum samples over 3 distributions

<u>الانجا</u>



Scheme 6: Performance

3 Unsatisfactory Performers (1 UK&I)

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	88 (25)	82 (25)	74 (25)	71 (23)	68 (23)	68 (23)	67 (24)
Number with Unsatisfactory Performance (< 80%) <mark>(UKal)</mark>	5 <mark>(D)</mark>	8 (0)	2 (U)	0 (U)	4 (0)	3 (D)	3 (1)
% Unsatisfactory Performance	5.7% (0%)	9.7% (0%)	2.7% (0%)	0% (0%)	5% (0%)	4.4% (D)	4.4% (4.1%)
38% neg 58% pos 4% samp	ative itive ples not asse	ssed					



Scheme 6: Unacceptable Performers 2024

Lab	Performance (<80%)	Issue
11	75.0%	Interpretation (pos cut off)
128	75.0%	Interpretation / kit issue
1418	58.3%	Interpretation / kit issue





Scheme 6: Kit Use and Performance



Manufacturer of Kit Used for Antibody Detection



OL Werfen Both Unknown

MANUFACTURER OF KIT USED FOR ANTIBODY DETECTION (N=68)



		Cla	ss l			Cla	ss II	
2024-25	One Lambda (n=34)	%	Werfen (n=19)	%	One Lambda (n=34)	%	Werfen (n=19)	%
601	Positive	100	Positive	100	Positive	100	Positive	100
602	Positive	100	Positive	100	Negative	84	Negative	100
603	Positive	100	Positive	100	Positive	100	Positive	100
604	Positive	67	Negative	80	Negative	100	Negative	87
605	Positive	100	Positive	100	Positive	100	Positive	100
606	Positive	100	Positive	94	Positive	100	Positive	100
607	Negative	94	Negative	100	Negative	94	Negative	94
608	Negative	84	Negative	100	Negative	77	Negative	94
609	Positive	100	Positive	94	Positive	100	Positive	100
610	Negative	66	Negative	100	Positive	100	Positive	100
611	Positive	100	Positive	100	Positive	97	Positive	100
612	Negative	97	Negative	94	Negative	97	Negative	100





HLA Antibody Specificity Analysis

Scheme 3: HLA Antibody Specificity Analysis

Purpose Assess participants ability to determine specificity of HLA antibodies

Satisfactory Performance 75% reports agree with consensus in distribution year Consensus At least 75% agreement on presence of HLA antibodies, 95% agreement on absense.

10 serum samples over 3 distributions

Scheme 3: Performance

Class I		2018	2019	2020	2021	2022	2023	2024	
Number of Participants (UK&I)		73 (25)	70 (25)	64 (24)	65 (24)	65 (24)	64 (24)	63 (23)	
Number with Jnsatisfactory Performance (UK&I)	Presence	15 (1)	3 (0)	1 (0)	1 (0)	1 (0)	4 (0)	2 (0)	
	Absence	5 (0)	2 (0)	1 (0)	1 (0)	1 (0)	1 (0)	3 (D)	
% Unsatisfactory Performance	Presence	20.5%	4.2%	1.6%	1.5%	1.5%	6.3%	3.2%	
	Absence	6.8%	2.6%	1.6%	1.5%	1.5%	1.5%	4.8%	
		Class	: 11	2018	2019	2020	202		
Uverall 3 labs with UP		Number of	nts (UK8	75 (25)	69 (25)	63 (24)	64 (24		
0 UKAI) -		Number wi	th	Pr	esence	12 (0)	5 (0)	2 (0)	3 (1







Class II		2018	2019	2020	2021	2022	2023	2024
Number of Participants (imber of Participants (UK&I)		69 (25)	63 (24)	64 (24)	64 (24)	64 (24)	63 (23)
Number with Unsatisfactory Performance (UK&I)	Presence	12 (0)	5 (0)	2 (0)	3 (0)	1 (0)	1(0)	2 (0)
	Absence	3 (D)	2 (D)	1 (0)	1 (0)	1 (0)	1 (0)	2 (0)
% Unsatisfactory Performance	Presence	16.0%	7.2%	3.2%	4.7%	1.6%	1.6%	3.2%
	Absence	4.0 %	2.8%	1.6%	1.6%	1.6%	1.6%	3.2%



Scheme 3: Unacceptable Performers 2024

3 labs (0 UK&I) with UP (<75%)

	Class I		Class	s II	САРА	Kit	
Lab	Presence	Absence	Presence	Absence			
302	62%	53%	57%	72%	No reply	Werfen	
1312	74%	55%	48%	52%	No reply	One Lambda	
1412		58%			Interpretation issue (pos cut off)	Werfen	

Scheme 3: Kit Use 2020-2024



Overall OL kits are the most widely used

UK&I labs are more likely to use a combination of kits

Werfen only kit use more prevalent in RoW labs

Scheme 3: Results by Kit Use



Similar percentage of antibodies reach consensus present (orange) in both kits

Less concordance in 'absent' antibodies

Greater percentage of Class I antibodies classed as not assessed in Werfen group



Scheme 3: Kit Use and Performance



Average overall satisfactory performance for detecting the 'presence' and 'absence' of antibodies was marginally higher for users of both kits.

2024-25

Avorago		Class I		Class II			
Performance	One Lambda	Werfen	Both	One Lambda	Werfen	Both	
	(n=34)	(n=14)	(n=12)	(n=34)	(n=14)	(n=12)	
Presence	94.1%	86.5%	97.2% 🕇	94.5%	89.7%	97.8% 🕇	
Absence	97.5%	94.2%	99.5% 🕇	97.4%	95.3%	99.5% 🕇	



Scheme 3: DQA/DPA Antibody Reporting

Reporting of antibodies to HLA-DQA and -DPA is optional and not assessed. Overall 42/63 (67%) report DQA, 38/63 (60%) report DPA

Previously 58% and 50%





An analysis of the data submitted for DQA and DPA antibodies in 2024-25 and 2023-24 was performed.

Large proportion of samples are negative or consensus absent.

No antibodies deemed positives.

Approx 25-30% not assessed.





HPA Antibody Detection/Specification


Scheme 11: HPA Antibody Detection/Specification

Purpose Assess participants ability to prrectly determine pesence and specificty of HPA antibodies.

Satisfactory Performance At least 75% of specificities in agreement with the consensus result in a distribution year.



Consensus

Presence of specificity determined by at least 75% agreement and absence determined by at least 95% agreement.

8 serum/plasma samples over 2 distributions

Scheme 11: Performance



• 1 Unsatisfactory Performers (0 UK&I)

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	35 (4)	39 (5)	42 (4)	43 (4)	43 (4)	43 (4)	44 (4)
Number with Unsatisfactory Performance (< 75%) (UKEI)	1 (0)	1 (0)	3 (0)	6 (0)	2 (0)	9 (0)	1 (0)
% Unsatisfactory Performance	2.9%	2.6%	7.1%	13.9%	4.5%	20.9%	2.3%

Scheme 11: HPA Antibody Detection/Specification

				HPA Antibody Consensus				
2024 Sample	HPA Detection	HLA Detection	Expected Result	Presence	Absence			
1	97.6% Pos	86.5% Pos	HPA-5b	HPA 5b 97.6%	HPA GP1a/11a 97.2%%			
2	100% Neg	94.4% Neg	HPA neg, HLA neg	N/A	N/A			
3	100% Pos	100% Neg	HPA-5b, HLA neg	HPA 5b 100%	HPA GP1a/11a 97.2%			
4	97.5% Pos	91.4% Neg	HPA-1a	HPA-1a 97.5%	HPA-3a 92.3%, 4b 90.9%, GPIIb/IIIa 91.7%			
5	100% Neg	94.3% Neg	HPA neg, HLA neg	N/A	N/A			
6	60% Neg	100% Pos	HPA-15b	N/A	HPA 15b 60% GP11b/111a 97.3%			
7	97.2% Neg	94.3% Neg	HPA neg, HLA neg	N/A	HPA 2b 97.2%			
8	97.1% Neg	94.3% Neg	HPA neg, HLA neg	N/A	HPA GP1b 97.1%			



Scheme 11: Unacceptable Performers 2024

Lab	HPA Presence	HPA Absence	Samples reported	Method	Error
1378	50%	100%	8/8	Immucor/Werfen PakLx	No response





Scheme 11: Analysis of Errors 2024

- Error rate extremely low (overall 0.15%) but errors often at clinically relevant polymorphisms.
- Errors found at HPA-1a (n=1, error rate 0.3%),
 2b (n=1, error rate 0.3%), 5b (n=1, error rate 0.3%) and some glycoproteins.

Errors 2024	HPA-1a	HPA-1b	HPA-2a	HPA-2b	HPA-3a	HPA-3b	HPA-4a	HPA-4b	HPA-5a	HPA-5b	HPA-6a	HPA-6b	HPA-15a	HPA-15b	GP IIb/IIIa	GP la/lla	GP Ib	GP Iv	CD109	Total
False Pos	0	0	0	1	0	0	0	0	0	0	0	0	0	0	1	2	1	0	0	5
False Neg	1	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	2
Total Errors	1	0	0	1	δ	0	0	0	C	1	0	0	0	0	1	2	1	0	0	7
% Error Rate	0.3	0.0	0.0	0.3	0.0	0.0	0.0	0.0	0.0	0.3	0.0	0.0	0.0	0.0	0.3	0.7	0.4	0.0	0.0	0.15
Total Tested	320	304	288	288	312	312	284	268	320	328	108	108	120	120	292	292	280	264	120	4728

• False positive (n=5) more common than false negative (n=2) errors.

Most labs had only 1 or 2 errors

Number	Number				
of Errors	of Labs				
1	3				
2	2				



Scheme 11: Selection of HPA Antibodies for Assessment



- Introduced in 2024-25
- Labs can select any/all HPA antibodies for assessment based on their clinical strategy
 Percentage of Participants Opting to Report at



Scheme 11: HPA-15 Detection

- 36% (15/42) of labs selected to be assessed for HPA-15 antibody detection
- We sent NIBSC Standard HPA-15b
- Sample 6/2024:

60% (9/15) reported HPA-15b absent 40% (6/15) reported HPA-15b present 100% reported HLA ab present





Key Data from the Schemes **Deborah Pritchard** UK NEQAS for H&I Director





HLA Phenotyping

Scheme 1A: HLA Phenotyping



Purpose

Assess participants ability to use serological and supplementary methods to correctly identify HLA phenotype

Satisfactory Performance 9 or more complete HLA phenotypes in agreement with consensus per distribution year.

10 blood samples over 5 distributions

At least 75% agreement on

Consensus

Scheme 1A: Performance



• 0 labs with unsatisfactory performance

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	38 (<mark>6</mark>)	38 (<mark>5</mark>)	34 (4)	33 (<mark>2</mark>)	28 (1)	23 (])	19 (0)
Number with Unsatisfactory Performance (< 90%) (UK&I)	6 (1)	8 (<mark>1</mark>)	3 (1)	2 (1)	2 (1)	2 (0)	0 (1)
% Unsatisfactory Performance	15.8%	21.1%	8.8%	6.1%	7.1%	8.7%	0%



Scheme 1A: 2024 Incorrect Assignments

3/190 (1.6%) incorrect HLA types in 2024 reported by 3 labs:

2 reports that contained broad not split specificity (e.g. B40 v B60)
1 clerical/typo error





Scheme 1A: 2024 Incorrect Assignments (not resulting in UPs)



Sample	ID	Consensus	Report
1A 03	268	A1, A26; B38, B57	A1, A26; B68 , B57
1A 05	147 + 159	A3, A23; B49, B65	A3, A23; B49, B14





Scheme



HLA Typing at 1st Field Resolution



Scheme 4A1: HLA Typing at 1st Field Resolution

Consensus

je za

At least 75% agreement on each allele. When consensus is not met, a reference result is used. Reference result is always used for DPB1 assessment

10 blood samples over 3 distributions

Purpose

Assess participants ability to correctly determine HLA genotypes at the 1st field resolution.

Satisfactory Performance 9 or more full HLA types in agreement with consensus/reference result in a distribution year.

Scheme 4A1: Performance



• 6 labs with unsatisfactory performance (2 UK&I)

	2018	2019	2020	2021	2022	2023	2024
Number of Participants	105	100	88	82	81	81	84
(UK&I)	(28)	(28)	(26)	(25)	(25)	(25)	(24)
Number with Unsatisfactory Performance (< 90%) (UKal)	15 (<mark>1</mark>)	4 (1)	8 (<mark>1</mark>)	6 (<mark>1</mark>)	7 (<mark>1</mark>)	11 (<mark>1</mark>)	6 (<mark>2</mark>)
% Unsatisfactory	14.3%	4%	9.1%	7.3%	8.4%	13.6%	7.1%
Performance	(3.6%)	(3.6%)	(0%)	(4%)	(0%)	(4%)	(8.3%)

Scheme 4A1: 2024-25 Incorrect Assignments



39/11493 (0.34%) errors reported by 20 different labs (3 UK&I) – last year 0.26%

29 samples contained an error:

- 21 samples with incorrect assignments *e.g. C*13 rather than C*07*
- 3 samples with nomenclature issue
- 3 missed assignment (reported homozygous when heterozygous)
- 2 samples with DRB3/4/5 presence/absence reported incorrectly

24 (83%) HLA types with one error 5 (17%) HLA types with multiple errors

> 14 (70%) labs made 1 error 3 (15%) labs made 2 errors 3 (15%) lab made 3 errors





Scheme 4A1: Unacceptable Performers 2024

Lab	Sample	Error	CAPA Response
41	01+02+03	Multiple issues	Transcription error / interpretation error
6	03+04+07	Incorrect DPB1* assignments	Limitations of kit
172	03+06+08	A* and DQB1* missed / assignment errors	Procedural error (new method)
1412	05+10	DQA1* and DPA1* and DPB1* missed assignments	Procedural error (reagent issue)
1418	07+09	Reporting errors	Transcription errors (staffing levels)
1443	09+10	Multiple reporting errors	EQA specific result entry errors





Interpretive HLA Genotype

Scheme 4A1: Interpretive HLA Genotype

Purpose Assess participants ability to correctly interpret their 4A1 genotype result to the 'split' specificity level.

Satisfactory Performance 9 or more full HLA types in agreement with consensus/reference result in a distribution year. Consensus

At least 75% agreement on each specificity. When consensus is not met, a reference result is used.

10 HLA genotypes from Scheme 4A1

Scheme 4A1i: Performance



5 lab with unsatisfactory performance (0 UK&I)

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	40 (21)	44 (22)	44 (22)	42 (21)	40 (21)	40 (21)	41 (20)
Number with Unsatisfactory Performance (< 90%) (UK21)	6 (1)	8 (<mark>1</mark>)	6 (<mark>2</mark>)	5 (1)	2 (1)	1(U)	5 (<mark>L</mark>)
% Unsatisfactory Performance	15.0%	18.1%	13.6%	11.9%	5.0%	2.5%	12.2%

24/5778 (0.42%) incorrect results reported by 7 different labs (0 UK&I) – last year 0.35%

• 17 samples contained an error:

Assignments

- 5 reporting at broad not split specificity level
- 3 samples with incorrect assignments
- 3 samples with missing assignment (reported homozygous when heterozygous)
- 3 sample with incorrect uses of nomenclature
- 3 samples with errors at presence/absence of DR51/52/53

Scheme 4A1i: 2024-25 Incorrect

12 (71%) HLA types with single errors 5 (29%) HLA type with multiple errors 2 (29%) labs made 1 error 5 (71%) labs made 2-4 errors









Scheme 4A1i: Unacceptable Performers 2024

Lab	Sample	Error	CAPA Response
1418	01-03 + 07	Broads instead of splits/multiple reporting errors	Staff training / interpretation
111	02+03+05	Incorrect nomenclature	EQA specific reporting errors (new participant)
1412	02+05	Multiple errors	Transcription errors
1372	02+06	Multiple errors	No response
1433	05+07	Multiple errors	Staff training / interpretation
	١		



Scheme 4A1: Types of Errors Over 5 Years





Scheme 4A1i: Serological Equivalents

- 4A1i Interpretative HLA Genotyping, which allows participants to translate genotypes to phenotypes
- Participants are expected to report to the 'split' specificity level using serological nomenclature, e.g. HLA-DQB1*03:01 should be reported as DQ7 (DQ3)
- Knowledge and exposure of phenotyping and converting between genotypes and phenotypes may no longer be so commonplace

Reliance on LIMS/analysis software

No longer perform phenotyping

- Errors common due to reporting issues (broad rather split or using incorrect nomenclature)
- CAPA repeatedly cite issues due to staff training and knowledge
 - Encourage utilisation of 4A1i for competency assessment

UK NEQAS have developed a template: https://ukneqashandi.org.uk/app/uploads/2025/03/Scheme-4A1i-Template.docx



Histocompatibility & Immunogenetics

UK NEQAS Education | Guality | Global Scheme 4A1 - Competency Assessment Templat

Task: Convert HLA genotypes reported in Scheme 4A1 to 'split specificity' level phenotypes (minimising computer aided assistance

Learning Objective: Accurately demonstrate knowledge of serological equivalents.



 Genetype
 A*
 A*
 B*
 B*
 C*
 C*
 DR*
 DR*
 DR83/4/5*
 DQ81*
 DQ81*
 Score

 Sample ID
 XX20XX
 Fhenetype
 A
 A
 B
 Bw4*
 Bw4*
 C
 C
 DR
 DR
 DR31/3/5*
 DQ81*
 DQ81*
 DQ81*
 Score

 Exploration
 Image: Score and Score and

**Each full phenotype (compared to consensus result on UK NEQAS Portal) deemed acceptable





Scheme



HLA Typing to 2nd or 3rd Field Resolution



Scheme 4A2: HLA Typing to 2nd or 3rd Field Resolution

Purpose Assess participants ability to correctly determine HLA type to 2nd or 3rd field.

Satisfactory Performance 9 or more full HLA types in agreement with consensus/reference genotype in a distribution year. Con At le allel rofer

Consensus

At least 75% agreement on each allele. If consensus is not met, a reference result is used.

10 blood samples over 3 distributions

Scheme 4A2: Performance

- 46/67 participants registered for 2nd field
- 22/67 participants registered for 3rd field
- 8 labs with unsatisfactory performance (1 UK&I)



	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UKAI)	63 (20)	62 (20)	64 (20)	63 (<mark>22</mark>)	61 (23)	65 (<mark>23</mark>)	67 (23)
Number with Unsatisfactory Performance (< 90%) (UK&I)	9 (<mark>2</mark>)	9 (1)	7 (0)	6 (1)	4 (1)	9 (<mark>2</mark>)	8 (1)
% Unsatisfactory Performance	14.3%	14.5%	11.0%	11.1%	6.5%	13.8% (8.7%)	11.9% (4.3%)

Scheme 4A2: Incorrect Assignments: 2nd Field



33/8846 (0.37%) incorrect HLA alleles reported by 12 labs (0 UK&I) – last year (0.55%)

- 14 reports of errors at the 2nd field
- e.g. DQA1*03:02 rather than DQA1*03:03
- 3 samples with alleles in a string that should have been resolved
- 3 reports of the wrong HLA type
- 3 reports of incorrect nomenclature
- e.g. the use of P / G groups

17 (74%) HLA types with a single error 6 (26%) HLA types with multiple errors



7 (58%) labs made 1 error 5 (42%) labs made 3-4 errors 19/3560 (0.53%) incorrect HLA alleles reported by 7 labs (1 UK&I) – last year (0.81%)

Scheme 4A2: Incorrect Assignments: 3rd Field

- 5 errors at 2nd field e.g. C*03:03:01 rather than 03:321:XX
- 5 incorrect assignments e.g. C*04:01:01 rather than C*05:01:01
- 3 errors at 3rd field e.g. DPB1*03:01:03 rather than DPB1*03:01:01
- 1 report of unresolved ambiguities e.g. reporting G groups

4 (29%) HLA types with multiple errors 10 (71%) HLA types with a single error





Number of Errors by Loci in 4A2 3rd Field 2024-25

4 (57%) labs made 1 error 2 (29%) labs made 2 errors 1 (14%) labs made 3+ errors

Scheme 4A2: Unacceptable Performers 2024

Lab	Sample	Error	Field	CAPA Response
284	01-03	Reported G groups	2 nd	EQA Specific Reporting issue
29	04+05	3rd Field: 4A2 04/2024: Reported B*40:01:01 homozygous, consensus B*40:01:02 homozygous 4A2 05/2024: Reported DPB1*03:01:03 , consensus DPB1*03:01:01	3 rd	Transcription error
112	02+05+08+10	2nd Field: 4A2 02/2024: Reported A*01:01, 01:03, consensus A*01:01 homozygous 4A2 05/2024: Reported C*03:03, consensus C*03:321 4A2 08/2024: Reported DQA1*03:02, consensus DQA1*03:03 4A2 10/2024: Reported DQA1*04:01, consensus DQA1*04:02	2 nd	Kit resolution issue
134	02+07+08+09	3rd Field: 4A2 02/2024: Reported DPB1*02:01:01, consensus DPB1*02:01:02 4A2 07/2024: Reported C*04:01:01, consensus C*05:01:01 4A2 08/2024: Reported DPB1*01:03:01, 02:01:02, consensus DPB1*01:01:01, 04:01:01 4A2 09/2024: Reported incomplete allele (C*07:02:0), consensus C*07:02:01, reported G groups for DRB1*	3 rd	No reply
309	03+07+08+09	3rd Field: 4A2 03/2024: Reported DRB3*01:01:02 homozygous, consensus DRB3*01:01:02, 03:01:01 4A2 07/2024: Reported B*15:01:01, consensus B*35:01:01 4A2 08/2024: Reported DRB1*01:01:01 and DQA1*03:01:01, consensus DRB1*03:01:01 and DQA1*03:03:01 4A2 09/2024: Reported B*07:01:01, consensus B*07:02:01	3 rd	Transcription error
1433	05+06+09	2nd Field: 4A2 05/2024: Reported C*03:03 , consensus C*03:321 4A2 06/2024: Reported B*42:02 , consensus B*44:02 4A2 09/2024: Reported DRB1*11:04 , consensus DRB1*11:01	2 nd	Transcription error / Kit resolution issue
185	08-10	2nd Field: 4A2 08/2024: Reported Unacceptable ambiguities DPB1*677:01/875:01N/1086:01 4A2 09/2024: Reported Unacceptable ambiguities DPB1*677:01/875:01N/1086:01 4A2 10/2024: Reported Unacceptable ambiguities DPB1*727:01/1285:01N, 677:01/875:01N/1086:01	2 nd	Unacceptable ambiguities / Kit resolution issue
223	07+09+10	2nd Field: 4A2 07/2024: Reported DRB1*15:02 , consensus DRB1*15:01 4A2 09/2024: Reported DQA1*04:01 , consensus DQA1*04:05 4A2 10/2024: Reported DQA1*04:01 , consensus DQA1*04:02	2 nd	Kit resolution issue



Scheme 4A2: Types of Errors Over 5 Years



Scheme 4A2 - Types of Error by Field (2020-24)





Scheme



KIR Genotyping



Scheme 9: KIR Genotyping

Purpose

Assess participants ability to correctly determine the presence or absence of specific KIR genes.

Satisfactory Performance 9 or more full KIR genotypes in agreement with consensus/reference genotype in a distribution year.





Consensus

At least 75% agreement on the presence/abesence of each gene. Reference type used where consensus is not met

10 blood samples over 2 distributions

Scheme 9: KIR Genotyping



- Participants able to report any of the following: KIR2DL1, KIR2DL2, KIR2DL3, KIR2DL4, KIR2DL5, KIR3DL1, KIR3DL2, KIR3DL3, KIR3DS1, KIR2DS1, KIR2DS2, KIR2DS3, KIR2DS4, KIR2DS5, KIR2DP1, KIR3DP1.
- Also able to report any other KIR polymorphisms they detected for information
- Participants can also report an 'A' or 'B' haplotype for each sample based on the gene content of the sample

Scheme 9: Performance



• 1 lab with unsatisfactory performance

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	9 (1)	12 (<mark>1</mark>)	12 (<mark>1</mark>)	15 (<mark>1</mark>)	15 (1)	15 (<mark>1</mark>)	13 (<mark>1</mark>)
Number with Unsatisfactory Performance (UK&I)	1 (<mark>0</mark>)	3 (<mark>0</mark>)	0 (0)	1(0)	0 (<mark>1</mark>)	1 (<mark>)</mark>	1(0)
% Unsatisfactory Performance	11.1%	25%	0%	6.7%	0%	6.7%	7.7%


Scheme 9: Unacceptable Performers 2024

Lab	Polymorphism	Error	CAPA Response
332	3DS1 & 2DS5	False Pos	Interpretation issues
	2DP1	False Neg	(staff training)







HPA Genotyping



Scheme 10: HPA Genotyping

Purpose Assess participants ability to correctly determine HPA polymorphisms.

Satisfactory Performance 9 or more full HPA types in agreement with consensus/reference genotype in a distribution year. Consensus At least 75% agreement on the presence/abesence of each allele. Reference type used where consensus is not met

je za

10 blood samples over 2 distributions

Scheme 10: HPA Genotyping



- Participants able to report any of the following: HPA-1, HPA-2, HPA-3, HPA-4, HPA-5, HPA-6, HPA-15
 - 35/38 reported HPA-1, 2, 3, 4, 5 and 15
 - 30/38 labs reported HPA-6
- Also able to report any other HPA polymorphisms detected, *for information*



Scheme 10: HPA Genotyping



• 0 lab with unsatisfactory performance

	2017	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK <mark>&</mark> I)	15 (5)	37 (6)	38 (6)	40 (U)	38 (6)	39 (6)	39 (<mark>6</mark>)	38 (<mark>6</mark>)
Number with Unsatisfactory Performance (< 100%) (UK&I)	1 (1)	1 (1)	3 (🛛)	0 (1)	0 (1)	1 (1)	1(1)	0 (0)
% Unsatisfactory Performance	6.7%	2.7%	7.9%	0%	0%	2.6%	2.6%	0%

Scheme 10: Errors in HPA Genotypes 2024

- 3 labs made 1 error
- Error rate extremely low 0.08% but errors at some clinically relevant polymorphisms.
- Errors found at HPA-3b (n=2), HPA-1b (n=1), HPA-4b (n=1)

Errors 2024	HPA-1 a	HPA-1 b	HPA-2 a	HPA-2 b	HPA-3 a	HPA-3 b	HPA-4 a	HPA-4 b	HPA-5 a	HPA-5 b	HPA-6 a	d ð-AqH	HPA-15 a	HPA-15 b	Total
False Neg	0	1	0	0	0	0	0	1	0	0	0	0	0	0	2
False Pos	0	0	0	0	0	2	0	0	0	0	0	0	0	0	2
Total Error	0	1	0	0	0	2	0	1	0	0	0	0	0	0	4
% Error	0.0	0.3	0.0	0.0	0.0	0.5	0.0	0.3	0.0	0.0	0.0	0.0	0.0	0.0	0.08
Total															
Tested	375	375	375	375	375	375	345	345	375	375	295	295	375	375	5030



• Even split of false positive (n=2) and false negative (n=2) errors.





HLA-B27 Testing

Scheme 1B: HLA-B27 Testing

Purpose Assess participants ability to correctly determine HLA-B27/2708/B*27 status.

Satisfactory Performance Making 10/10 reports that are in agreement with consensus in a distribution year.



Consensus At least 75% agreement on B27 status. Reference type used where consensus is not met

10 blood samples sent over 5 distributions

Scheme 1B: Performance



16 labs with unsatisfactory performance (3 UK&I)

	2018	2019	2020	2021	2022	2023	2024
Number of Participants	133	133	141	141	139	134	131
(UK&I)	(54)	(53)	(52)	(50)	(49)	(50)	(49)
Number with Unsatisfactory Performance (< 100%) (UK&)	10 (3)	4 (1)	12 (2)	3 (0)	8 (I)	11 (3)	16 (<mark>3</mark>)
% Unsatisfactory	7.5%	3.0%	8.5%	2.1%	5.7%	8.2%	12.2%
Performance (UK&I)	(5.5%)	(1.9%)	(3.8%)	(0%)	(0%)	(6%)	(6%)

5/10 samples distributed were HLA-B27 positive

Scheme 1B: 2024 Incorrect Assignments

Sample	Result	Lab Number	Technique	HLA Type	Lab dentified Caus
1B 02	False neg	1392 1435	Molecular Serological	B27 B44	No reply No reply
1B 03	False neg	40	Serological	B27 B40	Procedural/processing errors
1B 05 & 06	No results	1441	Serological	B27 B40 B27 B40	No reply
1B 06	False neg	1435 31 225 357	Serological Serological Molecular Serological	B27 B40	No reply Technical/testing issue Kit/interpretation/reagent issue No reply
1B 07 & 08	False pos/false neg	32	Molecular	B*27:08 B35 B7 B40	Sample mix up
1B 07	False neg	1431 106 324 409 1308 1312	Molecular Serological Serological Serological Serological Serological	B*27:08 B35	Reporting error Interpretation/reagent issue Interpretation/cut off values Interpretation issues No reply No reply
1B 09	False pos	324	Serological	B7 B35	Interpretation/cut off values
1B 09 & 10	No results	1402	Unknown	B7 B35 B7 B15	No reply
	• 88% false negat 70.5% errors inv Overall 73% use	ive, 12% false pos olved serological t molecular method	sitive techniques ds, 27% use serologi	cal methods	7/11 labs with unsatisfactory performance completed CAPA



Scheme



HFE Typing

Scheme 5A: HFE Testing

Purpose

Assess participants ability to correctly determine HFE mutations.

3 mutations assessed:

Codon 63: Histidine63Aspartic acid (H63D) Codon 282: cysteine282tyrosine (C282Y) Codon 65: Serine63Cysteine (S65C)

Satisfactory Performance

10 reports in agreement with consensus/reference result in a distribution year.





Consensus

At least 75% agreement on each HFE mutation. Reference type used where consensus is not met

10 donor samples sent over 3 distributions

Scheme 5A: Performance

3 labs with unsatisfactory performance (1 UK&I)





CAPA responses (n=2/3)

- Sample mix up 33.3%
- Transcription error 33.3%
- No reply 33.3%

Scheme

$\overline{7}$

HLA-B*57:01 Typing for Drug Hypersensitivity

Scheme 7: HLA-B*57:01 Typing for Drug Hypersensitivity

 \mathcal{Q}

Purpose

Q

Consensus

N S C

At least 75% agreement on the status of HLA-B*57:01. Reference result used when consensus not met.

10 blood samples over 3 distributions

correctly determine HLA-B*57:01 status

Satisfactory Performance Making 10 sample reports in agreement with the consensus/reference result in a distribution year.

Scheme 7: Performance



• 1 lab with unacceptable performance

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	67	67	67	64	52	50	46
	(27)	(27)	(27)	(25)	(18)	(18)	(18)
Number with Unacceptable	2	0	2	1	3	2	1
Performance (< 100%) (UK&I)	(0)	(0)	(0)	(1)	(<mark>0</mark>)	(U)	(0)
% Unsatisfactory Performance	3.0%	0.0%	3.1%	1.6%	5.8%	4.0%	2.2%

5/10 samples distributed were HLA-B*57:01 positive



Scheme 7: Unacceptable Performers 2024

Lab	Sample	Error	CAPA Response
308	03	False neg	No reply



New for 2025-26



Scheme 7: UPDATE

HLA-B*57:01 Typing for Drug Hypersensitivity

HLA-related Pharmacogenetics

- Abacavir Hypersensitivity **B*57:01**
- Allopurinol Hypersensitivity B*58:01
- Carbamazepine Hypersensitivity A*31:01, B*15:02
 - Oxcarbazepine Hypersensitivity B*15:02
 - Lamotrigine Hypersensitivity B*15:02
 - Flucloxacillin Hypersensitivity B*57:01
 - Phenytoin Hypersensitivity B*15:02
 - Tebentafusp Suitability A*02:01

NEW ASSESSMENT OPTIONS

(NO ADDITIONAL CHARGES)



All scheme 7 material will now be previously frozen whole blood



HLA Genotyping for Coeliac and other HLA Associated Disease

Scheme 8: HLA Genotyping for Coeliac and other HLA Associated Disease.

Purpose

Assess participants ability to correctly determine HLA type associated with various diseases e.g. coeliac disease, narcolepsy.

Satisfactory Performance

Making 10 sample reports in agreement with the reference genotype in a distribution year.

N S

Assessment

Lab results reported in format identical to clinical report. Reference HLA result used for assessment.

10 blood samples over 3 distributions

Scheme 8: Performance



10 Unsatisfactory Performers (1 UKEI)

	2018	2019	2020	2021	2022	2023	2024
Number of Participants (UK&I)	52	50	55	55	54	57	55
	(10)	(11)	(12)	(10)	(11)	(11)	(11)
Number with Unsatisfactory Performance (< 100%) (UK&I)	14 (4)	13 (2)	17 (5)	12 (2)	25 (5)	18 (2)	10 (1)
% Unsatisfactory Performance	27%	26%	31%	22%	46.3%	31.6%	18.2%
	(40%)	(18%)	(42%)	(20%)	(45%)	(18%)	(9.1%)

CAPA responses (n=8/10)

- Transcription errors 46%
- Kit interpretation error 17%
- Reporting error 7%
- Procedural error 7%
- Unknown 23%

Scheme 8: Unacceptable Performance by Disease

Disease	HLA Association	Number of Participants	No. of Participants with Unacceptable Performance
Coeliac	DQ2.5, DQ8, DQ2.2	51	9 (18%)
Narcolepsy	DQB1*06:02	24	0
Actinic Prurigo	DRB1*04:07	5	0
Birdshot Retinopathy	A*29	14	0
Behçet's	B*51	21	0
Rheumatoid Arthritis	DRB1*04	6	0
Diabetes	DR3, DR4	7	1 (14%)
Psoriasis	C*06	6	0
Allopurinol Hypersensitivity	B*58	8	0
Carbamazepine	A*31:01	9	0
Phenytoin	B*15:02	3	0
Tebentafusp	Tebentafusp A*02:01		0

New for 2025-26



Scheme 8: UPDATE HLA Genotyping for Coeliac and other HLA Associated Diseases

Coeliac Disease Narcolepsy Actinic Prurigo Birdshot Retinopathy Behcet's Disease Rheumatoid Arthritis Diabetes Psoriasis

Allopurinol hypersensitivity Carbamazepine hypersensitivity Phenytoin hypersensitivity Tebentafusp suitability

NOW PART OF SCHEME 7

Scheme 8: Interpretative Comments

• Interpretation of the genotype in terms of predisposition to CD not currently assessed

iv. DQ2.5 heterozygous (cis or trans)

HLA genotype result: HLA-DQA1*05, DQB1*02 HLA-DQA1*X, DQB1*X OR HLA-DQA1*05, DQB1*X HLA-DQA1*X, DQB1*02 HLA-DQB1*02, DQA1*05 (DQ2.5) positive HLA-DQB1*02, DQA1*02 (DQ2.2) negative HLA-DQB1*03:02 (DQ8) negative

Genotype comment: Positive for DQ2.5 (heterozygous)

Interpretative comment < 1 > : This individual has a genotype which is associated with coeliac disease

Interpretative comment < 2 > : This presence of DQA1*05, DQB1*02 (HLA-DQ2.5) has a strong association with coeliac disease in patients where laboratory tests or symptoms or endoscopic features suggest coeliac disease.



Pritchard D, Anand A, De'Ath A, Lee H, Rees MT. UK NEQAS and BSHI guideline: Laboratory testing and clinical interpretation of HLA genotyping results supporting the diagnosis of coeliac disease. Int J Immunogenet. 2024 Jan;51 Suppl 1:3-20. doi: 10.1111/iji.12649. Epub 2023 Dec 28. PMID: 38153308.

Scheme 8: Assessment of Interpretative Comments

• Pilot assessment based on points:

Coeliac Disease	Outcome	Improvement Point	Assessment
	HLA genotype aligned to reference type	N/A	Acceptable
HLA Genotype	Result not reported	N/A	Unacceptable
	HLA genotype not aligned to reference type	N/A	Unacceptable
	*HLA Comments / Correct Nomenclature Used	0	Acceptable
	*Incorrect HLA Comments / Incorrect Nomenclature Used	1	Unacceptable
Interpretation (>1	Risk of CD Present/Absent Correctly Identified	0	Acceptable
improvement	Risk of CD Present/Absent Incorrectly Identified	1	Unacceptable
point =	*Stratification of Risk Identified	0	Acceptable
unacceptable)	*Stratification of Risk Incorrectly Identified	1	Unacceptable
	Diagnostic Disclaimer Applied Correctly	0	Acceptable
	Diagnostic Disclaimer Not Applied or Incorrect	0.5	Acceptable

Pritchard D, Anand A, De'Ath A, Lee H, Rees MT. UK NEQAS and BSHI guideline: Laboratory testing and clinical interpretation of HLA genotyping results supporting the diagnosis of coeliac disease. Int J Immunogenet. 2024 Jan;51 Suppl 1:3-20. doi: 10.1111/iji.12649. Epub 2023 Dec 28. PMID: 38153308.

Scheme 8: Coeliac Disease – examples DRB1*07.01, 15:01; DQB1*02:02; 06:02; DQA1*01:02; 02:01 (DQ2.2, DQ6.1)



The presence of an associated HLA genotype does not confer a diagnosis of coeliac disease and has a low positive predictive value for coeliac disease.



Scheme Summary

Performance Summary for all Schemes

5 Year Trends in Unsatisfactory Performance



UK NEQAS for H&I Educational Crossmatch Scenario (EDXM) **Amy De'Ath** UK NEQAS for H&I Manager







"Schemes should relate more closely to clinical scenarios rather than testing individual test assays."

Whole Process 'EQA'





Assessed Schemes

- 1A, 4A1, 4A2 HLA Typing
- 6 HLA Antibody Detection
- 3 HLA Antibody Specification
- 2A, 2B Crossmatching

Educational Schemes

- Interpretative Educational Scenarios
- Educational Crossmatch Scheme
 - Clinical decision making based on results from multiple assays
 - Each assay only gives part of the picture
 - Results from one assay can influence the interpretation of another
 - Variation between centres (repertoires, cut-offs)



Educational Scheme Distribution



2024 Submissions

- 33 participants submitted results
- Not all labs reported results for all tests

• HLA genotype:

 \bigcirc

Consensus	A *	B *	C *	DRB1*	DRB4*	DQA1*	DQB1*	DPA1*	DPB1*
HLA Type	02:01	40:01	03:04	04:04	01:03	03:01	03:02	01:03	04:02
	32:01	51:01	15:02	09:01		03:02	03:03	02:01	11:01
Number of reports	32	32	32	32	24	29	32	21	29
% Labs in consensus	100%	100%	100%	100%	100%	100%	100%	100%	100%







Serum 1 Results

	Result	% Consensus	Comments			
HLA Class I Antibodies	No Consensus	58% (19/33)				
HLA Class II Antibodies	Negative	76% (25/33)				
DSA	None	100% (31/31)				
CDC XM	PBL Negative T cell Negative B cell No Consensus	100% (4/4) 100% (10/10) 71% (5/7 Neg)	CDC Negative – B cell XM with DTT 100% negative, without			
FCXM T Cell	Negative	96% (27/28)	FCXM Negative			
FCXM B Cell	Negative	96% (25/26)				
Transplant Risk	Low/Standard Intermediate	97% (29/30) 3% (1/30)				
Immunological Advice	Suitable for direct transplantation. Low level HLA antibodies present but not donor specific.					
Recommendations	Proceed to transplant.					


Serum 2 Results

 \bigcirc

	Result	% Consensus	Comments				
HLA Class I Antibodies	Positive	100% (33/33)	Multiple A, B and Cw ab >10,000				
HLA Class II Antibodies	Negative	85% (28/33)					
DSA	Present	100% (31/31)	100% B antibody reported at 6-20,000 39% Cw antibody 600-1,750				
CDC XM	PBL Negative T cell Negative B cell Negative	75% (3/4) 100% (10/10) 86% (6/7)	CDCXM Negative				
FCXM T Cell	Positive	89% (25/28)	FCXM Positive				
FCXM B Cell	Positive	81% (21/26)					
Transplant Risk	Intermediate High/Contraindication	16% (5/31) 84% (26/31)					
Immunological Advice	Not suitable for direct transplantation. High risk of AMR. If transplant proceeds use enhanced immunosuppression and post-transplant monitoring. Test for non-HLA and autologous antibodies.						
Recommendations	Seek alternative donor. Consider de-sensitisation. Monitor antibodies over time to consider de-listing. Discuss risk with patient.						





Serum 3 Results

 \bigcirc

	Result	% Consensus	Comments			
HLA Class I Antibodies	Positive	100% (33/33)	B and Cw antibodies 2,000 - >10,000 MFI			
HLA Class II Antibodies	Positive	100% (33/33)	DQ and DP antibodies 2,000 - >10,000 MFI			
DSA	Present	100% (31/31)	Multiple DSA to CI and CII ranging from 600-33,000			
CDC XM	PBL No consensus T cell Negative B cell Positive	50% (2/4 Pos) 78% (7/9) 86% (6/7)	CDCXM T cell Negative, B cell Positive FCXM Positive			
FCXM T Cell	Positive	100% (28/28)				
FCXM B Cell	Positive	92% (24/26)				
Transplant Risk	High/Contraindication	100% (31/31)				
Immunological Advice	Not suitable for direct transplantation. Risk of AMR. Consider de-sensitisation. If transplant proceeds use enhanced immunosuppression and post-transplant monitoring. Test for non-HLA and autologous antibodies.					
Recommendations	Seek alternative donor. Consider de-sensitisation. Monitor antibodies over time to consider de-listing. Discuss risk with patient.					

Summary of Crossmatch and DSA Detection Results

2024 Results		Serum 1		Serum 2		Serum 3	
DSA Defined by Luminex		Class I	Class II	Class I	Class II	Class I	Class II
MFI >10,000		N/A	N/A	B60 (100%)	N/A	B51 (100%) Cw10 (97%)	DR4 (100%)
MFI 5,001-9,999		N/A	N/A	N/A	N/A	Cw15 (84%)	DQ8 (97%) DQ9 (97%)
MFI 2,501-5,000		N/A	N/A	N/A	N/A	N/A	DR9 (45%) DQA1*03:02 (3%)
MFI <2,500		N/A	N/A	Cw10 (39%) Cw3 (3%)	N/A	N/A	DR53 (10%)
:XM ≡LL	No DTT	Negative		Negative		Positive	
DTT		Negative		Negative		Positive	
XM	T Cell	Negative		Positive		Positive	
FC	B Cell	Negative		Positive		Positive	
Risk		Low (97%) Intermediate (3%)		High (84%) Intermediate (16%)		High (100%)	



The table shows the percentage of participants identifying a DSA and the most common MFI range it was reported in.



Benefits





Benchmarking

Monitor performance of multiple techniques Make clinical interpretations on own results Compare local policies for clinical assessment Education

Monitor concordances Review variations Staff training

Competency

Laboratory staff Clinical staff



Future Considerations



Do you have any questions?

 \bigcirc

UKNEQASHandl@Wales.NHS.UK +44(0)1443 622185 www.ukneqashandi.org.uk